


MAD Nasal


Intranasal Mucosal Atomization Device

Using the MAD Nasal Intranasal Mucosal Atomization Device*


Procedure




1. Remove and discard the green vial adapter cap.




2. Pierce the medication vial with the syringe vial adapter.




3. Aspirate the proper volume of medication required to treat the patient (including medication to account for the dead space in the device).




4. Remove (twist off) the syringe from the vial adapter.




5. Attach the MAD Nasal Device to the syringe via the luer lock connector.



6. Using the free hand to hold the occiput of the head stable, place the tip of the MAD Nasal Device snugly against the nostril aiming slightly up and outward (toward the top of the ear).



7. Briskly compress the syringe plunger to deliver half of the medication into the nostril.



8. Move the device over to the opposite nostril and, repeating steps 6 and 7, administer the remaining medication into the nostril if indicated.

Tips to Improve Success

1. Minimise volume, maximise concentration.

- Use the appropriately concentrated drug

2. Maximise total mucosal absorptive surface area.

- Atomise the drug (rather than drip it in) to cover broad surface area
- Use BOTH nostrils to double the absorptive surface area
- Aim slightly up and outwards to cover the turbinates and olfactory mucosa

3. Beware of abnormal mucosal characteristics.

- Mucous, blood and vasoconstrictors reduce absorption
- Suction nostrils or consider alternate drug delivery method in these situations

MAD Nasal Intranasal Mucosal Atomization Device Specifications

Typical Particle Size	30–100 microns
System Dead Space	MAD100/MAD110/MAD130/MAD140 = 0.15 ml MAD300 = 0.06 ml
Tip Diameter	0.17" (4.3 mm)



MAD Nasal Device with vial adapter and 1 ml (or 3 ml) syringe

To learn more, visit teleflex.com

*For use with drugs approval for intranasal delivery

Teleflex, the Teleflex logo, MAD and MAD Nasal are trademarks or registered trademarks of Teleflex Incorporated or its affiliates, in the U.S. and/or other countries. Information in this document is not a substitute for the product Instructions for Use. The products in this document may not be available in all countries. Please contact your local representative. Subject to technical changes without further notice. © 2022 Teleflex Incorporated. All rights reserved.

MCI-100029-EN · REV 2 · 03 22 PDF

Distributed by:
Teleflex Headquarters International, Ireland · Teleflex Medical Europe Ltd. · IDA Business & Technology Park
 Dublin Road · Athlone · Co Westmeath · Tel. +353 (0)9 06 46 08 00 · Fax +353 (0)14 37 07 73 · orders.intl@teleflex.com
United Kingdom Tel. +44 (0)14 94 53 27 61 · info.uk@teleflex.com
South Africa Tel. +27 (0)11 807 4887 · assist.africa@teleflex.com

Teleflex